




Injection Date : 21 Jul 20
 Sample Name : A-769662-02

Method Info : Mobile Phase : A:water(0.01%TFA) B:ACN(0.01%TFA)
 Gradient:5%B increase to 95%B within 9min,95%B for 6min
 Flow Rate :1ml/min
 Column :Sunfire C18 150*4.6mm,3.5um
 Column Temperature:45C
 Detection:UV(214nm,4nm) and MS(ESI, POS Mode,110-1300amu)



Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
7.94	0.04	75.48	30.38	1.51
9.18	0.10	4939.61	627.02	98.49

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
7.94	0.04	40.06	16.86	0.90
9.18	0.10	4388.87	558.17	99.10