




LC/MS Report

Injection Date : 16 Mar 21
 Sample Name : SR1078-01
 Method Info : Mobile Phase : A:water(0.01%TFA) B:ACN(0.01%TFA)
 Gradient:5%B increase to 95%B within 1.5min,95%B for 1.5min
 Flow Rate :2.0ml/min
 Column :SunFire C18 3.5um, 4.6mm*50mm
 Column Temperature:50C
 Detection:UV(214nm,4nm) and MS(ESI, POS Mode,110-1300amu)



Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
1.85	0.01	1049.29	1237.72	98.06
2.09	0.01	20.71	23.26	1.94

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.85	0.01	753.29	891.45	100.00

Ret. Time: 1.85

