

File D:\DATA

Injection Date : 2 Mar 15 4:45

Sample Name : LDN-212854-01

Acq. Method : D:\METHODS\1-POS-3MIN.M

Analysis Method : D:\METHODS\1-POS-3MIN.M

Sample Info : *** No target masses specified ***

Method Info : Mobile Phase: A: water(0.01%TFA) B: ACN(0.01%TFA)

Gradient: 5%-95% B in 1.2min


Flow Rate: 2.2 ml/min

Column: Poroshell 120 EC-C18, 4.6*30mm, 2.7um


Oven Temperature: 50 C

Mass Range: 110-1500


*ADC1 A, ELSD




*DAD1 A, Sig=214,4 Ref=off



*DAD1 B, Sig=254,4 Ref=off



MSD1 TIC, MS File ES-API, Pos, Scan, Frag: 70



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
1.12	0.02	51.70	34.67	100.00

Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
1.12	0.02	2947.05	2025.27	100.00

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
1.12	0.02	2342.80	1597.16	99.13
1.26	0.03	20.54	8.27	0.87

LC/MS Report

Ret. Time: 1.12

<<<< POSITIVE SPECTRA >>>>

