





Injection Date : 25 Sep 20  
 Sample Name : 145108-58-3  
 Method Info : Mobile Phase : A:water(0.01%TFA) B:ACN(0.01%TFA)  
 Gradient:5% increase to 95%B within 1.3min,95%B for 1.7min  
 Flow Rate :2ml/min  
 Column :Sunfire, 50\*4.6mm,3.5um  
 Column Temperature:50C  
 Detection:UV(214,4nm) and MS(ESI,Pos mode,110 to 1000 amu)



Integration Results for ADC1 A, ELSD

RetTim	Width	Area	Height	Area%
1.44	0.02	117.83	94.00	100.00

Integration Results for DAD1 A, Sig=2144 Ref=off

RetTim	Width	Area	Height	Area%
1.44	0.04	4780.78	2096.73	100.00

Integration Results for DAD1 B, Sig=2544 Ref=off

RetTim	Width	Area	Height	Area%
1.44	0.03	83.40	43.40	100.00

Ret. Time: 1.44 <<<< POSITIVE SPECTRA >>>>

