




Injection Date : 17 Sep 20
 Sample Name : MSDC 0160-01
 Method Info : Column:SunFire C18 (4.6x 50 mm,3.5 um)
 Mobile phase: H2O(0.01%TFA) (A) / ACN(0.01%TFA)(B)
 Elution program: Gradient from 10 to95% of B in 1.4 min at 2ml/min
 Temperature: 50 °C
 Detection: UV (214 , 4 nm) and MS (ESI, Pos mode ,110 to 1000 amu)



Integration Results for DAD1 A, Sig=214,4 Ref=off


| RetTim | Width | Area | Height | Area% |
|--------|-------|--------|--------|--------|
| 1.82 | 0.01 | 222.25 | 277.98 | 100.00 |

Integration Results for DAD1 B, Sig=254,4 Ref=off

| RetTim | Width | Area | Height | Area% |
|--------|-------|--------|--------|--------|
| 1.82 | 0.01 | 194.40 | 242.82 | 100.00 |

Ret. Time: 1.39

<<<< POSITIVE SPECTRA >>>>



Ret. Time: 1.82

