




LC/MS Report

Injection Date : 30 Mar 21
Sample Name : RG108-03

Method Info : Mobile Phase : A:water(0.01%TFA) B:ACN(0.01%TFA)
Gradient:5%B increase to 95%B within 1.5min,95%B for 1.5min
Flow Rate :2.0ml/min
Column :SunFire C18 3.5um, 4.6mm*50mm
Column Temperature:50C
Detection:UV(214nm,4nm) and MS(ESI, POS Mode,110-1300amu)




Integration Results for DAD1 A, Sig=2144 Ref=off

RetTim	Width	Area	Height	Area%
1.52	0.02	4212.58	3374.53	100.00

Integration Results for DAD1 B, Sig=2544 Ref=off

RetTim	Width	Area	Height	Area%
1.52	0.01	625.20	692.73	100.00

Ret. Time: 1.23 <<<< POSITIVE SPECTRA >>>>



Ret. Time: 1.52

