



File D:\DATA\TIC10-HPLC

Acq. Method : D: HPLC
Analysis Method : D: HPLC
Sample Info : Easy-Access Method:
Method Info : Mobile Phase: A: Water(0.01%TFA) B: ACN(0.01%TFA)
Gradient: 5%B increase to 95%B within 8min,
95%B for7min,back to 5%B within 0.01min.
Flow Rate: 1 mL/min
Column: SunFire C18, 4.6*150mm,3.5um
Column Temperature: 40 C



Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
6.04	0.05	3318.52	1011.81	100.00

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
6.04	0.05	839.07	261.48	100.00