

File D:\DATA\TH-302-PRO3004-HPLC

Acq. Method : D: 1-HPLC.M

Analysis Method : D: 1-HPLC.M

Sample Info : Easy-Access Method: '1-HPLC'


Method Info : Mobile Phase: A: Water(0.01%TFA) B: ACN(0.01%TFA)
Gradient: 5%B increase to 95%B within 8min,
95%B for 7min, back to 5%B within 0.01min.

Flow Rate: 1 mL/min


Column: SunFire C18, 4.6*150mm, 3.5um

Column Temperature: 40 C

DAD1 A, Sig=214,4 Ref=off



DAD1 B, Sig=254,4 Ref=off



Integration Results for DAD1 A, Sig=214,4 Ref=off

RetTim	Width	Area	Height	Area%
7.19	0.04	889.50	330.54	100.00

Integration Results for DAD1 B, Sig=254,4 Ref=off

RetTim	Width	Area	Height	Area%
7.19	0.04	323.19	117.79	100.00